


NOTES: UNLESS OTHERWISE NOTED

1. MACHINED SURFACE ROUGHNESS 64  
O-RING GROOVES AND SURFACES 32
2. REMOVE ALL BURRS/SHARP CORNERS

BILL OF MATERIAL

NO.	REQ'D	REF DRW	DESCRIPTION	MATERIAL
N/A	1	SD4-023A	STRAP TO IMAGINGARRAY	COPPER OFHC

STAP TO BE  
MADE BY LAB-  
TECHNICIAN,  
STRAP WIDTH  
IS 1 INCH:  
NUMBER OF STRAPS  
TO BE FIGURED  
FROM A/L RATIO,  
SEE SD4-001B



REVDATE: N/A	REVISION: N/A
DATE: 04/09/99	<b>INSTITUTE FOR ASTRONOMY</b> 2680 WOODLAWN DR. HONOLULU, HI 96822
DESIGN: STAHLBERGER	
DRAWN: STAHLBERGER	
APPR'D: J. RAYNER	
<b>TOLERANCES</b>	
X.X	+/- .015
X.XX	+/- .005
X.XXX	+/- .002
ANGLE	+/- 30 min
FRACTION	+/- 1/32
INSTRUMENT: SPEX - CROSS DISPERSED SPECTROGRAPH	
PART NAME: STRAP TO IMAGINGARRAY	
TELESCOPE: IRTF	
DRW NO: SD4-023A	SCALE: APPROX 2
SIZE: A	JOS NO: 725
SHEET: 1 OF 1	

SD4-001B  
NEXT ASSY

PLOT: 10/19/99